

CE Documentation Review



No. 4U200319F.BGC0096

Holder:

BEIFA GROUP CO., LTD

No.68, Weiliu Road, Beilun District, Ningbo, Zhejiang, 315821, China

Review goal:

Verification of the presence of the Technical File in regards of the Medical Devices Directive 93/42/EEC Annex VII

Product:

Disposable medical mask (**no sterile**)

Model(s):

KZ001, KZ002

Classification:

Class I (no sterile)

(accordingly to the Manufacturer's declaration)

Review output:

We attest that a Technical File in reference to the Directive 93/42/EEC is in place for the CE Marking process.

The manufacturer is responsible for the CE Marking process, and not exempted to carry out all necessary compliance activities. This document has been issued on the basis of the regulation on ECM Voluntary Mark for the certification of products. RG01_ECM rev.3 available at: www.entecerma.it

Issuance date: 19 March 2020

Expiry date: 18 March 2025

Reviewer
Technical expert
Amanda Payne

Approver
ECM Service Director
Luca Bedonni

Ente Certificazione Macchine

Via Cà Bella, 243 - 40053 Valsamoggia Loc. Castello di Serravalle (Bo) Italy
☎ +39.0516705141 📠 +39.0516705156 ✉ info@entecerma.it 🌐 www.entecerma.it

Test Report

Customer : BEIFA GROUP CO., LTD
Address : NO.68, WEILIU ROAD, BEILUN DISTRICT, NINGBO, ZHEJIANG, 315821
Received Date : Mar 03, 2020
Turn Around Time : Mar 03, 2020 to Mar 24, 2020
Sample Description : MASK
Item No. : KZ001, KZ002
Buyer : N/A
Manufacturer : BEIFA GROUP CO., LTD

<u>Test Specification</u>	<u>Conclusion</u>
EN 14683:2019 Medical face masks – Requirements and test methods, Type II and Type IIR	Pass
Bacterial filtration efficiency (BFE)	Pass
Breathability (Differential Pressure)	Pass
Microbial Cleanliness	Pass
Synthetic Blood Penetration Resistance	Pass

***** FOR FURTHER DETAILS, PLEASE REFER TO THE FOLLOWING PAGE(S) *****

Laboratory Contact Information


Technology and Complaints : Hedda@newpath.com.cn | +86.17798587371
Billing and Inquiry : Hedi.Ammar@newpath.com.cn | +86.17798587376


Authorized By :




Amy Wang / Authorized Signatory

Sample Photo



Tested Component(s)			
	/	/	/
1.Mask	2.	3.	4.

Abbreviation

ND = Not Detected (less than RL).

RL = Reporting Limit.

NT = Not Tested.

N/A = Not Applicable.

mg/kg = milligram per kilogram = ppm.

1 mg/kg = 0.0001%.

NM = Not Meet.

NC = No Comment.

R1 = Revised report with Synthetic Blood Penetration Resistance test results.

Test Result

Bacterial filtration efficiency (BFE)

Test Method : With reference to EN 14683:2019, the BFE test is performed to determine the filtration efficiency of test articles by comparing the bacterial control counts upstream of the test article to the bacterial counts downstream. A suspension of Staphylococcus aureus was aerosolized using a nebulizer and delivered to the test article at a constant flow rate and fixed air pressure. The challenge delivery was maintained at $1.7 - 2.7 \times 10^3$ colony forming units (CFU) with a mean particle size (MPS) of $3.0 \pm 0.3 \mu\text{m}$. The aerosols were drawn through a six-stage, viable particle, Andersen sampler for collection. This test method complies

Test Side : Inside

Test Area : 40 cm^2

Flow Rate : 28 Liters per minute (L/min)

Test Condition : $85 \pm 5\%$ relative humidity (RH) and $21 \pm 5^\circ\text{C}$ for a minimum of 4 hours

Sample Dimension : 90mm x 165mm

Positive Control : 2,000 CFU

Negative Monitor : <1 CFU

MPS : $2.7 \mu\text{m}$

Test Result :

<u>Parameter</u>	<u>1</u>	<u>2</u>	<u>3</u>	<u>4</u>	<u>5</u>
Percent BFE (%)	99.4	99.4	99.2	99.0	99.5
Limit	≥ 98	≥ 98	≥ 98	≥ 98	≥ 98
Conclusion	Pass	Pass	Pass	Pass	Pass

Breathability (Differential Pressure)

Test Method : With reference to EN 14683:2019, The Differential Pressure (Delta P) test is performed to determine the breathability of test articles by measuring the differential air pressure on either side of the test article using a manometer, at a constant flow rate.

Test Side : Inside

Test Area : 40 cm^2

Flow Rate : 8 Liters per minute (L/min)

Test Condition : $85 \pm 5\%$ relative humidity (RH) and $21 \pm 5^\circ\text{C}$ for a minimum of 4 hours

Sample Dimension : 90mm x 165mm

Test Result :

<u>Parameter</u>	<u>1</u>	<u>2</u>	<u>3</u>	<u>4</u>	<u>5</u>
Delta P (Pa/cm ²)	17.2	17.0	17.5	17.0	17.3
Limit	<60	<60	<60	<60	<60
Conclusion	Pass	Pass	Pass	Pass	Pass

Microbial Cleanliness

- Test Method : With reference to EN 14683:2019, The testing was conducted in accordance with EN 14683:2014, with the exception of approximate volumes of eluent used when performing the extraction procedure and a temperature range of 30-35°C used for aerobic incubation.
- When bioburden results are calculated using a software program, manual calculations may differ slightly due to rounding. The counts determined on products are colony forming units and may not always reflect individual microorganisms. The sponsor performs any statistical analysis and determines the acceptable limits.
- Controls/Monitors : Bacillus atrophaeus
- Extract Fluid : Peptone Tween salt
- Fluid Volume : 300 mL
- Extract Method : Orbital Shaking for 5 minutes at 250 rpm
- Plating Method : Membrane Filtration
- Agar Medium : Tryptic Soy Agar
Sabouraud Dextrose Agar with Chloramphenicol
- Recovery Efficiency : Exhaustive Rinse Method
- Aerobic Bacteria : Plates were incubated 3 days at 30-35°C, then enumerated
- Fungal : Plates were incubated 7 days at 20-25°C, then enumerated
- Test Result :

Parameter	<u>1</u>	<u>2</u>	<u>3</u>	<u>4</u>	<u>5</u>
Weight (g)	3.1	3.1	3.1	3.1	3.1
Aerobic	<3	<3	<3	<3	<3
Fungal	<3	<3	<3	<3	<3
Total Bioburden (CFU/mask)	<6.0	<5.9	<6.0	<6.0	<5.9
Total Bioburden (CFU/g)	<2	<2	<2	<2	<2
Limit	<30 CFU/g	<30 CFU/g	<30 CFU/g	<30 CFU/g	<30 CFU/g
Conclusion	Pass	Pass	Pass	Pass	Pass

Synthetic Blood Penetration Resistance

- Test Method : This procedure was performed to evaluate surgical facemasks and other types of protective clothing materials designed to protect against fluid penetration. The purpose of this procedure is to simulate an arterial spray and evaluate the effectiveness of the test article in protecting the user from possible exposure to blood and other body fluids. The distance from the target area surface to the tip of the cannula is 30.5 cm. A test volume of 2 mL of synthetic blood was employed using the targeting plate method.
- This test method was designed to comply with ASTM F1862 and ISO 22609 (as referenced in EN 14683:2019 and AS4381:2015) with the following exception: ISO 22609 requires testing to be performed in an environment with a temperature of 21 ± 5°C and a relative humidity of 85 ± 10%. Instead, testing was performed at ambient conditions within one minute of removal from the environmental chamber held at those parameters.
- Tested Number : 32 pcs
- Test Side : Outside
- Pre-Conditioning : Minimum of 4 hours at 21 ± 5°C and 85 ± 5% relative humidity (RH)
- Test Conditions : 19.2°C and 32% RH

Test Pressure : 120 mmHg (16.0 kPa)

Test Result :

<u>Parameter</u>	<u>1-7, 9-21, 23-27, 29-32</u>	<u>8</u>	<u>22</u>	<u>28</u>
Synthetic Blood Penetration	None Seen	Yes	Yes	Yes

Per ASTM F1862 and ISO 22609, an acceptable quality limit of 4.0% is met for a normal single sampling plan when ≥ 29 of 32 tested articles show passing results.

Conclusion**Pass*******End of Report*****

This report is made solely on the basis of your instructions and/or information and materials supplied by you. It is not intended to be a recommendation for any particular course of action. Newpath does not accept a duty of care or any other responsibility to any person other than the Client in respect of this report and only accepts liability to the Client insofar as is expressly contained in the terms and conditions governing Newpath's provision of services to you. Newpath makes no warranties or representations either express or implied with respect to this report save as provided for in those terms and conditions. We have aimed to conduct the Review on a diligent and careful basis and we do not accept any liability to you for any loss arising out of or in connection with this report, in contract, tort, by statute or otherwise, except in the event of our gross negligence or wilful misconduct.

This report shall not be reproduced except in full, without written approval of the laboratory.

This statement of conformity is only based on the actual measured test result by the laboratory, without taking the influence of uncertainty into account.